


Intoxilyzer 9000 Evidential Breath Alcohol Test (EBAT) Test Re

Subject Name: MISTICH, ANTHONY DOB: 08/31/1976 Sex: Male Driver License: 120720232 State: CO Zip Code: 80227	Location: CDPHE EVIDENTIAL BREATH ALCOHOL TESTING PROGRAM Serial Number: 90-000210 Test Number: 0210140003 Case Number: FREE TEXT Date & Time of Offense: 01/10/2014 10:28
Intoxilyzer Operator: MISTICH, ANTHONY J Last Certification Date: 11/15/2013 Agency: COLORADO DEPT OF PUBLIC HEALTH	Arresting Officer: MISTICH, ANTHONY J Agency: COLORADO DEPT OF PUBLIC HEALTH Crash: Yes Injury: No

MISTICH, ANTHONY
Test Date: 01/10/2014

20-minute deprivation period	Start Time: 11:05	End Time: 11:25
<input checked="" type="checkbox"/> The 20-minute deprivation period has been conducted at the certified EBAT facility by a certified EBAT instructor or operator that is in an active status.	<input checked="" type="checkbox"/> The subject has been deprived access to foreign material that may be introduced into the mouth cavity during the 20-minute deprivation period.	
<input checked="" type="checkbox"/> The subject has removed any foreign material from the mouth cavity that is not permanent in nature, prior to starting the 20-minute deprivation period.	<input checked="" type="checkbox"/> The subject has been observed for signs of belching, regurgitation, or intake of any foreign material into the mouth cavity during the 20-minute deprivation period.	

EBAT Test Sequence	g/210L	Time
Air Blank	0.000	11:25:25
Sim Temp	33.9 C	
Wet Cal Chk	0.101	11:25:52
Air Blank	0.000	11:26:32
Subject Sample	0.143	11:27:29
Breath Volume	2.60 Liters	
Air Blank	0.000	11:28:35
Diagnostics	Passed	11:28:39
Air Blank	0.000	11:29:16
Wait		11:31:20
Air Blank	0.000	11:31:57
Subject Sample	0.144	11:33:30
Breath Volume	2.43 Liters	
Air Blank	0.000	11:34:21
Sim Temp	33.9 C	
Wet Cal Chk	0.098	11:34:48
Air Blank	0.000	11:35:28
Final Result	0.143 g/210L	


COMMENTS:
DEMO EBAT

I am currently certified as an EBAT operator and attest that this Evidential Breath Alcohol Test (EBAT) has been performed in compliance (5CCR 1005-2) following the procedures set forth by the CDPHE.

I authenticate the information provided on this report is produced by the certified EBAT instrument used to perform this test.

01/10/2014

Certified EBAT Instructor or Operator Signature

Date

COLORADO DEPARTMENT OF
PUBLIC HEALTH AND ENVIRONMENT

Pursuant to the Colorado Board of Health Rules Pertaining To The Testing Of Alcohol And Other Drugs (5-CCR1005-2), The CDPHE certifies and approves the use of the listed Intoxilyzer 9000 (I-9000) to perform Evidential Breath Alcohol Testing (EBAT) for the purpose of determining alcohol content

I-9000 Serial Number

90-000210

Certification Period

04/05/2013 - 10/13/2016

unless previously rescinded.



Colorado Department
of Public Health
and Environment

Laura Gillim-Ross

Laura Gillim-Ross, PhD, Director
Laboratory Services Division

Certificate # 0210130002I

Intoxilyzer 9000 (I-9000) Instrument Performance Report (IPR)

Subject Name:
 MISTICH, ANTHONY
 DOB: 08/31/1976
 Sex: Male
 Driver License 120720232
 State: CO
 Zip Code: 80227

Location: CDPHE EVIDENTIAL
 BREATH ALCOHOL
 TESTING PROGRAM
 Serial Number: 90-000210
 Test Number: 0210140003
 Case Number: FREE TEXT
 Date & Time of Offense: 01/10/2014 10:28

I-9000 Certified Solution Change Record

Test Date: 06/19/2013
 Test Time: 14:46:45
 Instructor User ID: 467497
 Lot Number: 201208D
 Bottle Number: 255
 Expiration Date: 08/07/2014
 Target Value: 0.100 g/210L
 Target Range: (0.095 - 0.105)g/210L

Test	g/210L	Time
Air Blank	0.000	14:47:08
Wet Cal Chk	0.098	14:47:35
Air Blank	0.000	14:48:14
Wet Cal Chk	0.098	14:48:41
Air Blank	0.000	14:49:21
Wet Cal Chk	0.097	14:49:47
Air Blank	0.000	14:50:27
Wet Cal Chk	0.099	14:50:53
Air Blank	0.000	14:51:33
Wet Cal Chk	0.099	14:52:00
Air Blank	0.000	14:52:39
Wet Cal Chk	0.098	14:53:06
Air Blank	0.000	14:53:47
Wet Cal Chk	0.098	14:54:14
Air Blank	0.000	14:54:53
Wet Cal Chk	0.098	14:55:20
Air Blank	0.000	14:56:00
Wet Cal Chk	0.098	14:56:26
Air Blank	0.000	14:57:06
Wet Cal Chk	0.098	14:57:32
Air Blank	0.000	14:58:12
Average	0.0981	
Standard Deviation	0.0006	
Relative Std Dev	0.0058	
60-day Average	0.097	

I-9000 Verification Record

Test Date: 04/05/2013
 Test Time: 11:04:57
 Technician User ID: 467497
 Certificate #: 0210130002I

Alcohol Value	Alcohol Found	Acceptable Range
.020	0.018	(0.015 - 0.025)
.040	0.040	(0.035 - 0.045)
.050	0.049	(0.045 - 0.055)
.080	0.075	(0.075 - 0.085)
.100	0.098	(0.095 - 0.105)
.150	0.146	(0.143 - 0.158)
.200	0.199	(0.190 - 0.210)
.300	0.297	(0.285 - 0.315)
.400	0.401	(0.380 - 0.420)
Interferent	*	

Verification Test: PASS Invalid Sample: PASS
 Stability Test: PASS Improper Sample: PASS
 EBAT Test: PASS

I-9000 60-Day Exception Message Record

Exception Message	Number
* Invalid Sample	0
* No 0.020 Agreement	0
* Range Exceeded	0
* Test Sequence Aborted	0
* Subject Refused	0
* Deficient Sample	0
* Calibration Check Out Of Tolerance	0
* Diagnostic Failure	0
* Unstable Reference	0
* Simulator Communication Failure	0
* Calibration Check Correlation Failure	0
* Purge Failure	0
* RFI Detected	0
* Ambient Failure	0
* Improper Sample	0
* Interferent Detected	0
* Simulator Temperature Out Of Range	0
* Simulator Exception Encountered	0
Exception Message Total	0
Evidential Breath Alcohol Test Total	2

I-9000 Calibration Adjustment Record

Date - 04/04/2013 Technician User ID 467497

I-9000 Certification Record

Certification Period
 From: 04/05/2013 To: 04/05/2014
 From: To:
 From: To:

Intoxilyzer 9000 (I-9000) Instrument Performance Report (IPR)

Subject Name:

MISTICH, ANTHONY

DOB: 08/31/1976

Sex: Male

Driver License 120720232

State: CO.

Zip Code: 80227

Location: CDPHE EVIDENTIAL

BREATH ALCOHOL

TESTING PROGRAM

Serial Number: 90-000210

Test Number: 0210140003

Case Number: FREE TEXT

Date & Time of Offense: 01/10/2014 10:28

I-9000 Maintenance History Record

<u>Date</u>	<u>Technician Name</u>	<u>Description</u>
10/13/2015	1157044	I-9000 CALIBRATION VERIFICATION TESTING WAS SUCCESSFULLY COMPLETED. THIS INSTRUMENT IS CERTIFIED BY CDPHE TO PERFORM EVIDENTIAL BREATH ALCOHOL TESTING.
11/06/2014	731869	I-9000 CALIBRATION VERIFICATION TESTING WAS SUCCESSFULLY COMPLETED. THIS INSTRUMENT IS CERTIFIED BY CDPHE TO PERFORM EVIDENTIAL BREATH ALCOHOL TESTING.
09/23/2014	467497	TEST DATA ENTRY LINE 2, LINE 3
04/28/2014	731869	I-9000 CALIBRATION VERIFICATION TESTING WAS SUCCESSFULLY COMPLETED. THIS INSTRUMENT IS CERTIFIED BY CDPHE TO PERFORM EVIDENTIAL BREATH ALCOHOL TESTING.
03/01/2013	467497	THIS IS A TEST DATA ENTRY
04/05/2013	467497	I-9000 CALIBRATION VERIFICATION TESTING WAS SUCCECSSLFULLY COMPLETED. THIS INSTRUMENT IS CERTIFIED BY CDPHE TO PERFORM EVIDENTIAL BREATH ALCOHOL TESTING.
04/04/2013	467497	I-9000 CALIBRATION ADJUSTMENT SUCCESSFULLY COMPLETED